

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102820\
 Data File : VX019161.D
 Acq On : 28 Oct 2020 13:32
 Operator : JC/MD
 Sample : L4470-04DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW179D1-GW-20201019DL

Quant Time: Oct 29 04:40:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	327621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	567125	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	528059	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	265315	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	214741	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
35) Dibromofluoromethane	5.46	113	172519	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
50) Toluene-d8	8.70	98	684481	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.12	95	263294	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	1199	0.378	ug/l	88
27) cis-1,2-Dichloroethene	4.56	96	29813	7.300	ug/l	90
44) Trichloroethene	7.19	130	237816	55.924	ug/l	100
45) 1,2-Dichloropropane	7.49	63	1017	0.280	ug/l #	85
64) Tetrachloroethene	9.32	164	20764	5.161	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102820\
Data File : VX019161.D
Acq On : 28 Oct 2020 13:32
Operator : JC/MD
Sample : L4470-04DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW179D1-GW-20201019DL

Quant Time: Oct 29 04:40:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

