

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102821\
 Data File : VX025005.D
 Acq On : 28 Oct 2021 14:41
 Operator : JC/MD
 Sample : VX1028MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK618

Manual Integrations
 APPROVED

MMDadoda
 11/1/2021 12:58:48 PM

Quant Time: Oct 29 06:48:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 28 12:38:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	147435	50.000	ug/L	# 0.00
28) Chlorobenzene-d5	10.055	117	133576	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	56069	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	50246	53.464	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.920%		
7) Chloroethane-d5	1.672	69	36983	57.570	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 115.140%		
11) 1,1-Dichloroethene-d2	2.306	63	91874	39.672	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 79.340%		
21) 2-Butanone-d5	4.470	46	106220	113.153	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 113.150%		
24) Chloroform-d	5.068	84	130408	53.030	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 106.060%		
26) 1,2-Dichloroethane-d4	5.964	65	97953	54.318	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.640%		
32) Benzene-d6	5.976	84	204496	53.937	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.880%		
36) 1,2-Dichloropropane-d6	7.311	67	71136	55.056	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 110.120%		
41) Toluene-d8	8.653	98	189349	52.202	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 104.400%		
43) trans-1,3-Dichloroprop...	8.951	79	39708	53.477	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 106.960%		
47) 2-Hexanone-d5	9.384	63	75566m	114.869	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 114.870%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	96395	55.172	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 110.340%		
66) 1,2-Dichlorobenzene-d4	12.323	152	61912	56.905	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 113.800%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102821\
Data File : VX025005.D
Acq On : 28 Oct 2021 14:41
Operator : JC/MD
Sample : VX1028MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK618

Manual Integrations
APPROVED

MMDadoda
11/1/2021 12:58:48 PM

Quant Time: Oct 29 06:48:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 28 12:38:37 2021
Response via : Initial Calibration

