

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032452.D
 Acq On : 28 Oct 2022 12:19
 Operator : JC/MD
 Sample : N5335-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1

Quant Time: Oct 29 04:31:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203932	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83952	48.625	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.240%
35) Dibromofluoromethane	5.385	113	67144	46.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	243154	48.071	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.140%
62) 4-Bromofluorobenzene	11.079	95	90192	45.647	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.300%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	1849	2.947	ug/l	91
17) Carbon Disulfide	2.514	76	1061	0.357	ug/l #	92
30) Chloroform	5.092	83	43357	15.988	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032452.D
Acq On : 28 Oct 2022 12:19
Operator : JC/MD
Sample : N5335-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1

Quant Time: Oct 29 04:31:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

