

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032457.D
 Acq On : 28 Oct 2022 14:16
 Operator : JC/MD
 Sample : N5355-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5518-1

Quant Time: Oct 29 04:32:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206416	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84694	48.026	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.060%
35) Dibromofluoromethane	5.385	113	66411	45.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.200%
50) Toluene-d8	8.647	98	243771	47.613	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.220%
62) 4-Bromofluorobenzene	11.079	95	91532	45.768	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.540%
Target Compounds						
86) p-Isopropyltoluene	12.012	119	6397	1.120	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032457.D
Acq On : 28 Oct 2022 14:16
Operator : JC/MD
Sample : N5355-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5518-1

Quant Time: Oct 29 04:32:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

