

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032458.D
 Acq On : 28 Oct 2022 14:39
 Operator : JC/MD
 Sample : N5355-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5518-2

Quant Time: Oct 29 04:32:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203150	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82821	47.859	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.720%		
35) Dibromofluoromethane	5.391	113	66087	45.606	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.220%		
50) Toluene-d8	8.653	98	241405	47.909	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.820%		
62) 4-Bromofluorobenzene	11.079	95	91229	46.350	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.700%		
Target Compounds						
86) p-Isopropyltoluene	12.012	119	2517	0.437	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032458.D
Acq On : 28 Oct 2022 14:39
Operator : JC/MD
Sample : N5355-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5518-2

Quant Time: Oct 29 04:32:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

