

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032461.D
 Acq On : 28 Oct 2022 15:49
 Operator : JC/MD
 Sample : N5336-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-1

Quant Time: Oct 29 04:33:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	119125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81069	48.978	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.385	113	64927	45.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.560%
50) Toluene-d8	8.647	98	235538	47.758	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	87500	45.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.840%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	470	0.341	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032461.D
Acq On : 28 Oct 2022 15:49
Operator : JC/MD
Sample : N5336-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-1

Quant Time: Oct 29 04:33:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

