

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032469.D
 Acq On : 28 Oct 2022 18:55
 Operator : JC/MD
 Sample : N5193-08
 Misc : 9.07g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-SS-04-05-15ME

Quant Time: Oct 29 04:34:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82835	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85904	48.105	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.220%
35) Dibromofluoromethane	5.385	113	66778	44.274	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.540%
50) Toluene-d8	8.647	98	250589	47.780	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.560%
62) 4-Bromofluorobenzene	11.079	95	83864	40.935	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.880%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032469.D
Acq On : 28 Oct 2022 18:55
Operator : JC/MD
Sample : N5193-08
Misc : 9.07g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-SS-04-05-15ME

Quant Time: Oct 29 04:34:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

