

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032470.D
 Acq On : 28 Oct 2022 19:18
 Operator : JC/MD
 Sample : N5193-09ME
 Misc : 7.55g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-SS-04-15-3ME

Quant Time: Oct 29 04:34:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79835	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83344	47.274	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.540%
35) Dibromofluoromethane	5.385	113	63110	42.682	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.360%
50) Toluene-d8	8.647	98	243153	47.293	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.580%
62) 4-Bromofluorobenzene	11.085	95	79546	39.607	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	79.220%#
Target Compounds						
18) Methyl Acetate	2.709	43	3413	1.638	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032470.D
Acq On : 28 Oct 2022 19:18
Operator : JC/MD
Sample : N5193-09ME
Misc : 7.55g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-SS-04-15-3ME

Quant Time: Oct 29 04:34:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

