

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
 Data File : VX032471.D
 Acq On : 28 Oct 2022 19:41
 Operator : JC/MD
 Sample : N5193-11ME 10X
 Misc : 5.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-SS-06-20-40ME

Quant Time: Oct 29 04:34:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110035	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84321	47.830	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.660%
35) Dibromofluoromethane	5.385	113	65764	44.625	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.260%
50) Toluene-d8	8.653	98	262701	51.265	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.520%
62) 4-Bromofluorobenzene	11.079	95	124569	62.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	124.460%#
Target Compounds						
					Qvalue	
31) Cyclohexane	5.489	56	52590	24.603	ug/l #	83
39) Methylcyclohexane	7.385	83	307119	126.487	ug/l	96
67) Ethyl Benzene	10.195	91	14923	1.793	ug/l	96
73) Isopropylbenzene	10.963	105	438893	50.335	ug/l	99
78) n-propylbenzene	11.305	91	1843495	183.507	ug/l	99
83) tert-Butylbenzene	11.719	119	29216	3.927	ug/l	83
85) sec-Butylbenzene	11.890	105	458855	51.256	ug/l	91
89) n-Butylbenzene	12.335	91	1059191	160.275	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102822\
Data File : VX032471.D
Acq On : 28 Oct 2022 19:41
Operator : JC/MD
Sample : N5193-11ME 10X
Misc : 5.19g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-SS-06-20-40ME

Quant Time: Oct 29 04:34:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

