

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102823\
 Data File : VX038624.D
 Acq On : 28 Oct 2023 01:04
 Operator : JC/MD
 Sample : VX1028MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK637

Quant Time: Oct 28 04:45:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102823WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 28 02:45:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	443574	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	428429	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	230745	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	140000	46.752	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.500%
7) Chloroethane-d5	1.672	69	111885	48.120	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.240%
11) 1,1-Dichloroethene-d2	2.307	65	62377	46.835	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.660%
21) 2-Butanone-d5	4.452	46	231956	99.453	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.450%
24) Chloroform-d	5.056	84	281483	46.308	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.620%
26) 1,2-Dichloroethane-d4	5.958	65	182705	47.403	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.800%
32) Benzene-d6	5.977	84	578721	49.948	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.900%
36) 1,2-Dichloropropane-d6	7.306	67	169402	50.289	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.580%
41) Toluene-d8	8.647	98	558798	48.371	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.740%
43) trans-1,3-Dichloroprop...	8.952	79	78061	47.428	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.860%
47) 2-Hexanone-d5	9.385	63	199667	101.357	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.360%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	270281	49.212	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.420%
66) 1,2-Dichlorobenzene-d4	12.317	152	229830	50.728	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.460%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102823\
Data File : VX038624.D
Acq On : 28 Oct 2023 01:04
Operator : JC/MD
Sample : VX1028MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK637

Quant Time: Oct 28 04:45:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102823WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 28 02:45:59 2023
Response via : Initial Calibration

