

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048384.D
 Acq On : 28 Oct 2025 09:27
 Operator : JC/MD
 Sample : VX1028WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1028WBL01

Quant Time: Oct 29 01:13:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	143516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	286237	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	281396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129160	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	113622	56.467	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.940%
35) Dibromofluoromethane	5.373	113	79707	40.789	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	81.580%
50) Toluene-d8	8.634	98	324329	47.351	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.700%
62) 4-Bromofluorobenzene	11.061	95	128432	49.289	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048384.D
Acq On : 28 Oct 2025 09:27
Operator : JC/MD
Sample : VX1028WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1028WBL01

Quant Time: Oct 29 01:13:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

