

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048389.D
 Acq On : 28 Oct 2025 11:32
 Operator : JC/MD
 Sample : Q3460-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-2

Quant Time: Oct 29 01:17:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	130820	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	265154	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	274876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	135974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	104025	56.714	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.420%
35) Dibromofluoromethane	5.373	113	73172	40.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	80.840%
50) Toluene-d8	8.635	98	298751	47.085	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.180%
62) 4-Bromofluorobenzene	11.061	95	127981	53.021	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.040%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	13051	20.741	ug/l	96
20) Methylene Chloride	2.788	84	19500	12.719	ug/l	98
43) Isopropyl Acetate	6.336	43	8815	2.516	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048389.D
Acq On : 28 Oct 2025 11:32
Operator : JC/MD
Sample : Q3460-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-2

Quant Time: Oct 29 01:17:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

