

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048392.D
 Acq On : 28 Oct 2025 12:33
 Operator : JC/MD
 Sample : Q3460-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-5

Quant Time: Oct 29 01:18:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound		R.T. QIon		Response	Conc Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene		5.544	168	119027	50.000 ug/l	0.00
34) 1,4-Difluorobenzene		6.751	114	246697	50.000 ug/l	0.00
63) Chlorobenzene-d5		10.037	117	257009	50.000 ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.006	152	126783	50.000 ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4		5.940	65	92133	55.208 ug/l	0.00
Spiked Amount 50.000		Range 78 - 117		Recovery	= 110.420%	
35) Dibromofluoromethane		5.373	113	67835	40.278 ug/l	0.00
Spiked Amount 50.000		Range 75 - 124		Recovery	= 80.560%	
50) Toluene-d8		8.635	98	275954	46.746 ug/l	0.00
Spiked Amount 50.000		Range 92 - 112		Recovery	= 93.500%	
62) 4-Bromofluorobenzene		11.061	95	117214	52.193 ug/l	0.00
Spiked Amount 50.000		Range 83 - 123		Recovery	= 104.380%	
Target Compounds						
						Qvalue
16) Acetone		2.374	43	13444	23.483 ug/l	95
20) Methylene Chloride		2.788	84	19592	14.045 ug/l	97
43) Isopropyl Acetate		6.324	43	9661	2.964 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048392.D
Acq On : 28 Oct 2025 12:33
Operator : JC/MD
Sample : Q3460-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-5

Quant Time: Oct 29 01:18:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

