

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048396.D
 Acq On : 28 Oct 2025 14:09
 Operator : JC/MD
 Sample : Q3472-06 20X
 Misc : 2.17g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-2

Quant Time: Oct 29 01:20:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	118728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	242873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	253038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131717	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	104858	62.991	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	125.980%#
35) Dibromofluoromethane	5.373	113	72485	43.716	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.440%
50) Toluene-d8	8.635	98	301509	51.879	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.760%
62) 4-Bromofluorobenzene	11.061	95	130604	59.071	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	118.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048396.D
Acq On : 28 Oct 2025 14:09
Operator : JC/MD
Sample : Q3472-06 20X
Misc : 2.17g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-2

Quant Time: Oct 29 01:20:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

