

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048397.D
 Acq On : 28 Oct 2025 14:30
 Operator : JC/MD
 Sample : Q3472-08 20X
 Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-3

Quant Time: Oct 29 01:20:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	107776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	215456	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	225138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	112893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	85911	56.853	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.700%
35) Dibromofluoromethane	5.373	113	63128	42.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.840%
50) Toluene-d8	8.635	98	266022	51.598	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.200%
62) 4-Bromofluorobenzene	11.061	95	114039	58.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048397.D
Acq On : 28 Oct 2025 14:30
Operator : JC/MD
Sample : Q3472-08 20X
Misc : 1.20g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-3

Quant Time: Oct 29 01:20:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

