

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
 Data File : VX048401.D
 Acq On : 28 Oct 2025 15:52
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 29 01:21:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	149859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	298949	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	309191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	159072	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	120737	57.463	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.920%
35) Dibromofluoromethane	5.379	113	88082	43.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.320%
50) Toluene-d8	8.634	98	373384	52.195	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.400%
62) 4-Bromofluorobenzene	11.067	95	160479	58.969	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102825\
Data File : VX048401.D
Acq On : 28 Oct 2025 15:52
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 29 01:21:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

