

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102918\
 Data File : VX005652.D
 Acq On : 29 Oct 2018 11:21
 Operator : JC/MD
 Sample : VX1029MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029MBL01

Quant Time: Oct 30 02:23:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	220581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303122	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	267367	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130104	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	127444	47.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.66%	
35) Dibromofluoromethane	5.50	113	96910	47.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.44%	
50) Toluene-d8	8.71	98	357197	48.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.86%	
62) 4-Bromofluorobenzene	11.14	95	119183	42.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102918\
Data File : VX005652.D
Acq On : 29 Oct 2018 11:21
Operator : JC/MD
Sample : VX1029MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1029MBL01

Quant Time: Oct 30 02:23:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

