

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102918\
 Data File : VX005664.D
 Acq On : 29 Oct 2018 16:57
 Operator : JC/MD
 Sample : J5718-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WS-CFA-111C

Quant Time: Oct 30 03:42:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218214	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	311711	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	278599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137154	51.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.98%	
35) Dibromofluoromethane	5.51	113	103516	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
50) Toluene-d8	8.71	98	368468	48.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.16%	
62) 4-Bromofluorobenzene	11.14	95	124925	43.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.24%	
Target Compounds						
16) Acetone	2.45	43	12520	8.895	ug/l	95
25) 2-Butanone	4.70	43	3050	1.423	ug/l #	49
52) Toluene	8.79	92	2159	0.410	ug/l	98
67) Ethyl Benzene	10.26	91	3344	0.343	ug/l	100
68) m/p-Xylenes	10.36	106	4478	1.182	ug/l	96
69) o-Xylene	10.70	106	4414	1.239	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	7376	1.023	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	9838	1.346	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102918\
Data File : VX005664.D
Acq On : 29 Oct 2018 16:57
Operator : JC/MD
Sample : J5718-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WS-CFA-111C

Quant Time: Oct 30 03:42:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

