

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102918\
 Data File : VX005669.D
 Acq On : 29 Oct 2018 19:11
 Operator : JC/MD
 Sample : J5692-05DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ171I-20181024DL

Quant Time: Oct 30 04:00:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	264535	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	397435	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	341329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174303	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153810	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	5.51	113	121353	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
50) Toluene-d8	8.71	98	447488	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
62) 4-Bromofluorobenzene	11.14	95	163992	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
6) Chloroethane	1.72	64	5529	2.647	ug/l #	88
12) 1,1-Dichloroethene	2.37	96	5294	2.132	ug/l	87
24) 1,1-Dichloroethane	3.70	63	149401	28.394	ug/l	99
32) 1,1,1-Trichloroethane	5.49	97	14265	3.284	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102918\
Data File : VX005669.D
Acq On : 29 Oct 2018 19:11
Operator : JC/MD
Sample : J5692-05DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ171I-20181024DL

Quant Time: Oct 30 04:00:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

