

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102918\
 Data File : VX005671.D
 Acq On : 29 Oct 2018 20:04
 Operator : JC/MD
 Sample : J5692-03RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ172I-20181024RE

Quant Time: Oct 30 04:14:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	275536	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	411844	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	344871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	168121	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154995	46.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.16%	
35) Dibromofluoromethane	5.50	113	123842	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
50) Toluene-d8	8.71	98	462064	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
62) 4-Bromofluorobenzene	11.14	95	157226	41.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.10%	
Target Compounds						
6) Chloroethane	1.73	64	16300	7.492	ug/l	98
12) 1,1-Dichloroethene	2.37	96	10271	3.971	ug/l	93
24) 1,1-Dichloroethane	3.70	63	344510	62.861	ug/l	100
32) 1,1,1-Trichloroethane	5.50	97	11995	2.651	ug/l	99
85) sec-Butylbenzene	11.94	105	4937	0.472	ug/l	84
93) 1,2,4-Trichlorobenzene	13.65	180	6042	1.530	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102918\
Data File : VX005671.D
Acq On : 29 Oct 2018 20:04
Operator : JC/MD
Sample : J5692-03RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ172I-20181024RE

Quant Time: Oct 30 04:14:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

