

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102918\
 Data File : VX005677.D
 Acq On : 29 Oct 2018 23:37
 Operator : JC/MD
 Sample : VX1029WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1029WBL02

Quant Time: Oct 30 05:00:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	260653	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	377813	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	167824	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150285	47.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.46%	
35) Dibromofluoromethane	5.51	113	121482	47.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.98%	
50) Toluene-d8	8.71	98	442455	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
62) 4-Bromofluorobenzene	11.14	95	151720	43.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102918\
Data File : VX005677.D
Acq On : 29 Oct 2018 23:37
Operator : JC/MD
Sample : VX1029WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1029WBL02

Quant Time: Oct 30 05:00:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

