

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_X\DATA\VX102919\
 Data File : VX013274.D
 Acq On : 29 Oct 2019 17:26
 Operator : JC/SP
 Sample : K5602-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190937-FIELD-BLANK

Quant Time: Oct 30 03:38:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	67201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	96052	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	95913	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	59960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	74247	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
35) Dibromofluoromethane	5.48	113	57250	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
50) Toluene-d8	8.71	98	198406	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	11.13	95	73225	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
Target Compounds						
20) Methylene Chloride	2.84	84	1627	1.595	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_X\DATA\VX102919\
Data File : VX013274.D
Acq On : 29 Oct 2019 17:26
Operator : JC/SP
Sample : K5602-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190937-FIELD-BLANK

Quant Time: Oct 30 03:38:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

