

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
 Data File : VX005691.D
 Acq On : 30 Oct 2018 15:00
 Operator : JC/MD
 Sample : PB114242ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114242ZHE#02

Quant Time: Oct 31 01:29:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	240705	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	340416	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	313002	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151287	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	146620	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
35) Dibromofluoromethane	5.50	113	113076	49.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.12%	
50) Toluene-d8	8.71	98	410851	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.14	95	139290	44.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.06%	
Target Compounds						
8) Diethyl Ether	2.19	74	1853	1.12	ug/l	91
16) Acetone	2.45	43	11430	7.36	ug/l	96
20) Methylene Chloride	2.85	84	3239	1.25	ug/l #	91
43) Isopropyl Acetate	6.46	43	160331	25.38	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
Data File : VX005691.D
Acq On : 30 Oct 2018 15:00
Operator : JC/MD
Sample : PB114242ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114242ZHE#02

Quant Time: Oct 31 01:29:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

