

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
 Data File : VX005693.D
 Acq On : 30 Oct 2018 15:54
 Operator : JC/MD
 Sample : PB114242ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114242ZHE#04

Quant Time: Oct 31 01:47:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	346627	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	142826	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	142096	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	5.51	113	108743	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	8.71	98	396103	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
62) 4-Bromofluorobenzene	11.14	95	132895	41.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	2.19	74	2166	1.36	ug/l	96
16) Acetone	2.44	43	11749	7.86	ug/l	100
18) Methyl Acetate	2.78	43	9557	1.16	ug/l	96
20) Methylene Chloride	2.85	84	3728	1.49	ug/l	92
43) Isopropyl Acetate	6.46	43	155176	24.12	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
Data File : VX005693.D
Acq On : 30 Oct 2018 15:54
Operator : JC/MD
Sample : PB114242ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114242ZHE#04

Quant Time: Oct 31 01:47:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

