

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
 Data File : VX005696.D
 Acq On : 30 Oct 2018 17:14
 Operator : JC/MD
 Sample : J5194-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-05-102618-A

Quant Time: Oct 31 02:04:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	229469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	346385	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	300962	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146342	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	141374	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	5.50	113	108696	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
50) Toluene-d8	8.71	98	398519	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
62) 4-Bromofluorobenzene	11.14	95	133702	42.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.02%	
Target Compounds						
8) Diethyl Ether	2.19	74	3591	2.28	ug/l	96
16) Acetone	2.45	43	30675	20.72	ug/l	99
20) Methylene Chloride	2.86	84	5007	2.02	ug/l	93
25) 2-Butanone	4.70	43	3835	1.70	ug/l #	87
43) Isopropyl Acetate	6.46	43	153871	23.94	ug/l	98
95) Naphthalene	13.83	128	10908	1.09	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
Data File : VX005696.D
Acq On : 30 Oct 2018 17:14
Operator : JC/MD
Sample : J5194-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-05-102618-A

Quant Time: Oct 31 02:04:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

