

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
 Data File : VX005697.D
 Acq On : 30 Oct 2018 17:40
 Operator : JC/MD
 Sample : J5697-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CHRT-58-66-98-92

Quant Time: Oct 31 02:06:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	211520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	323359	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	278098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	134039	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136054	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
35) Dibromofluoromethane	5.50	113	103328	47.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.38%	
50) Toluene-d8	8.71	98	363458	46.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.40%	
62) 4-Bromofluorobenzene	11.14	95	123692	41.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.26%	

Target Compounds						Qvalue
16) Acetone	2.45	43	7106	5.21	ug/l	92
20) Methylene Chloride	2.85	84	2959	1.29	ug/l	# 86
43) Isopropyl Acetate	6.46	43	136309	22.72	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103018\
Data File : VX005697.D
Acq On : 30 Oct 2018 17:40
Operator : JC/MD
Sample : J5697-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CHRT-58-66-98-92

Quant Time: Oct 31 02:06:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

