

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013278.D
 Acq On : 30 Oct 2019 11:16
 Operator : JC/SP
 Sample : VX1030WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1030WBL01

Quant Time: Oct 31 01:56:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	81313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	114219	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	112333	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	69481	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	85140	50.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.12%	
35) Dibromofluoromethane	5.48	113	64522	50.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.32%	
50) Toluene-d8	8.71	98	225724	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.13	95	85768	45.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.82%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013278.D
Acq On : 30 Oct 2019 11:16
Operator : JC/SP
Sample : VX1030WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1030WBL01

Quant Time: Oct 31 01:56:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

