

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013283.D
 Acq On : 30 Oct 2019 14:03
 Operator : JC/SP
 Sample : PB124420TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB124420TB

Quant Time: Oct 31 02:18:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	70929	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	100573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	99118	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	61187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	68592	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
35) Dibromofluoromethane	5.48	113	51737	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
50) Toluene-d8	8.71	98	181262	43.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.22%	
62) 4-Bromofluorobenzene	11.13	95	70563	42.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.78%	
Target Compounds						
20) Methylene Chloride	2.84	84	1492	1.423	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013283.D
Acq On : 30 Oct 2019 14:03
Operator : JC/SP
Sample : PB124420TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB124420TB

Quant Time: Oct 31 02:18:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

