

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013285.D
 Acq On : 30 Oct 2019 14:49
 Operator : JC/SP
 Sample : K5602-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190942-AMENDED-COMPOSITE-

Quant Time: Oct 31 02:28:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	68840	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	102502	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	101899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	61217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	81618	56.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.38%	
35) Dibromofluoromethane	5.49	113	145	0.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.26%	
50) Toluene-d8	8.71	98	214379	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.13	95	74819	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
Target Compounds						
16) Acetone	2.44	43	8654	12.758	ug/l	99
20) Methylene Chloride	2.85	84	2671	2.382	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013285.D
Acq On : 30 Oct 2019 14:49
Operator : JC/SP
Sample : K5602-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190942-AMENDED-COMPOSITE-

Quant Time: Oct 31 02:28:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

