

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013286.D
 Acq On : 30 Oct 2019 15:12
 Operator : JC/SP
 Sample : K5603-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190949-AMND-COMPOSITE-KM-

Quant Time: Oct 31 02:32:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	68543	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	102049	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	101880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	62623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	77084	53.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.54%	
35) Dibromofluoromethane	5.50	113	56	0.05	ug/l	0.01
Spiked Amount 50.000			Recovery	=	0.10%	
50) Toluene-d8	8.71	98	201361	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	11.14	95	74829	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	
Target Compounds						
16) Acetone	2.43	43	6904	10.222	ug/l	96
20) Methylene Chloride	2.85	84	3897	3.356	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013286.D
Acq On : 30 Oct 2019 15:12
Operator : JC/SP
Sample : K5603-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190949-AMND-COMPOSITE-KM-

Quant Time: Oct 31 02:32:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

