

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013287.D
 Acq On : 30 Oct 2019 15:59
 Operator : JC/SP
 Sample : K5602-07RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190941-AMENDED-COMPOSITE-

Quant Time: Oct 31 02:35:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	72161	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	104541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	102594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	65929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	79142	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
35) Dibromofluoromethane	5.49	113	590	0.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	1.02%	
50) Toluene-d8	8.71	98	207934	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
62) 4-Bromofluorobenzene	11.14	95	77180	45.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.28%	
Target Compounds						
16) Acetone	2.43	43	10661	14.993	ug/l	93
20) Methylene Chloride	2.84	84	2183	1.921	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013287.D
Acq On : 30 Oct 2019 15:59
Operator : JC/SP
Sample : K5602-07RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190941-AMENDED-COMPOSITE-

Quant Time: Oct 31 02:35:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

