

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
 Data File : VX013289.D
 Acq On : 30 Oct 2019 16:45
 Operator : JC/SP
 Sample : K5603-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20190949-AMND-COMPOSITE-KM-

Quant Time: Oct 31 03:09:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	69876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	100336	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	100605	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	63541	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	77108	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
35) Dibromofluoromethane	5.49	113	97	0.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.18%	
50) Toluene-d8	8.71	98	204139	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.13	95	73447	44.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.50%	
Target Compounds						
16) Acetone	2.44	43	6853	9.953	ug/l	94
20) Methylene Chloride	2.85	84	3688	3.136	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX103019\
Data File : VX013289.D
Acq On : 30 Oct 2019 16:45
Operator : JC/SP
Sample : K5603-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20190949-AMND-COMPOSITE-KM-

Quant Time: Oct 31 03:09:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

