


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX103118\
Data File : VX005704.D
Acq On : 31 Oct 2018 11:56
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Title : SW846 8260
Last Update : Wed Oct 31 18:23:06 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	1598	PASS
75	95	30	60	57.0	4090	PASS
95	95	100	100	100.0	7173	PASS
96	95	5	9	7.7	550	PASS
173	174	0.00	2	1.8	101	PASS
174	95	50	100	80.0	5737	PASS
175	174	5	9	7.9	451	PASS
176	174	95	101	97.8	5611	PASS
177	176	5	9	7.0	395	PASS