


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX103120\
Data File : VX019188.D
Acq On : 29 Oct 2020 13:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Title : SW846 8260
Last Update : Tue Oct 27 13:55:07 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	12325	PASS
75	95	30	60	50.0	34802	PASS
95	95	100	100	100.0	69661	PASS
96	95	5	9	6.5	4551	PASS
173	174	0.00	2	0.8	446	PASS
174	95	50	100	78.0	54362	PASS
175	174	5	9	7.0	3793	PASS
176	174	95	101	95.9	52106	PASS
177	176	5	9	6.5	3404	PASS