

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103122\
 Data File : VX032482.D
 Acq On : 31 Oct 2022 13:00
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Nov 01 05:08:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135133	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218896	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90662	48.285	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.580%
35) Dibromofluoromethane	5.391	113	71841	46.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	262054	48.266	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.540%
62) 4-Bromofluorobenzene	11.079	95	102033	48.110	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103122\
Data File : VX032482.D
Acq On : 31 Oct 2022 13:00
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Nov 01 05:08:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

