

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103122\
 Data File : VX032493.D
 Acq On : 31 Oct 2022 17:17
 Operator : JC/MD
 Sample : N5340-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7-102722

Quant Time: Nov 01 05:12:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82948	47.506	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.020%
35) Dibromofluoromethane	5.385	113	66315	45.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.260%
50) Toluene-d8	8.647	98	238724	47.244	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.480%
62) 4-Bromofluorobenzene	11.079	95	92445	46.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.971	59	24444	96.184	ug/l #	94
16) Acetone	2.380	43	87041	137.156	ug/l	99
30) Chloroform	5.105	83	4619	1.684	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103122\
Data File : VX032493.D
Acq On : 31 Oct 2022 17:17
Operator : JC/MD
Sample : N5340-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7-102722

Quant Time: Nov 01 05:12:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

