

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX103122\
 Data File : VX032499.D
 Acq On : 31 Oct 2022 19:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 05:17:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.577	0.599	-3.8	91	0.00
3 P	Chloromethane	0.488	0.482	1.2	90	0.00
4 C	Vinyl Chloride	0.537	0.539	-0.4#	90	0.00
5 T	Bromomethane	0.429	0.500	-16.6	106	0.00
6 T	Chloroethane	0.360	0.350	2.8	96	0.00
7 T	Trichlorofluoromethane	1.017	1.020	-0.3	93	0.00
8 T	Diethyl Ether	0.330	0.328	0.6	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.536	0.2	95	0.00
10 T	Methyl Iodide	0.558	0.669	-19.9	104	0.00
11 T	Tert butyl alcohol	0.101	0.092	8.9	85	0.01
12 CM	1,1-Dichloroethene	0.501	0.495	1.2#	93	0.00
13 T	Acrolein	0.101	0.118	-16.8	114	0.00
14 T	Allyl chloride	0.799	0.798	0.1	93	0.00
15 T	Acrylonitrile	0.274	0.257	6.2	91	0.00
16 T	Acetone	0.253	0.227	10.3	91	0.00
17 T	Carbon Disulfide	1.194	1.149	3.8	87	0.00
18 T	Methyl Acetate	0.821	0.804	2.1	92	0.00
19 T	Methyl tert-butyl Ether	1.762	1.756	0.3	96	0.00
20 T	Methylene Chloride	0.579	0.552	4.7	96	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.535	2.0	94	0.00
22 T	Diisopropyl ether	1.729	1.711	1.0	96	0.00
23 T	Vinyl Acetate	1.397	1.376	1.5	92	0.00
24 P	1,1-Dichloroethane	0.994	0.985	0.9	94	0.00
25 T	2-Butanone	0.391	0.359	8.2	89	0.00
26 T	2,2-Dichloropropane	0.791	0.770	2.7	91	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.640	-1.4	95	0.00
28 T	Bromochloromethane	0.354	0.317	10.5	90	0.00
29 T	Tetrahydrofuran	0.249	0.232	6.8	89	0.00
30 C	Chloroform	1.091	1.109	-1.6#	98	0.00
31 T	Cyclohexane	0.842	0.841	0.1	92	0.00
32 T	1,1,1-Trichloroethane	0.979	1.011	-3.3	96	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.693	0.3	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.357	0.348	2.5	94	0.00
36 T	1,1-Dichloropropene	0.497	0.502	-1.0	95	0.00
37 T	Ethyl Acetate	0.517	0.474	8.3	93	0.00
38 T	Carbon Tetrachloride	0.551	0.565	-2.5	95	0.00
39 T	Methylcyclohexane	0.588	0.582	1.0	90	0.00
40 TM	Benzene	1.405	1.406	-0.1	94	0.00
41 T	Methacrylonitrile	0.263	0.268	-1.9	96	0.00
42 TM	1,2-Dichloroethane	0.575	0.587	-2.1	96	0.00
43 T	Isopropyl Acetate	0.794	0.788	0.8	92	0.00
44 TM	Trichloroethene	0.396	0.398	-0.5	96	0.00
45 C	1,2-Dichloropropane	0.361	0.363	-0.6#	96	0.00
46 T	Dibromomethane	0.277	0.278	-0.4	96	0.00
47 T	Bromodichloromethane	0.524	0.541	-3.2	97	0.00
48 T	Methyl methacrylate	0.410	0.399	2.7	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX103122\
 Data File : VX032499.D
 Acq On : 31 Oct 2022 19:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 05:17:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	88	0.01
50 S	Toluene-d8	1.240	1.232	0.6	95	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.489	2.2	91	0.00
52 CM	Toluene	0.917	0.930	-1.4#	95	0.00
53 T	t-1,3-Dichloropropene	0.535	0.559	-4.5	94	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.594	-3.8	95	0.00
55 T	1,1,2-Trichloroethane	0.386	0.383	0.8	96	0.00
56 T	Ethyl methacrylate	0.556	0.568	-2.2	92	0.00
57 T	1,3-Dichloropropane	0.624	0.621	0.5	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.264	1.5	90	0.00
59 T	2-Hexanone	0.371	0.362	2.4	90	0.00
60 T	Dibromochloromethane	0.422	0.437	-3.6	97	0.00
61 T	1,2-Dibromoethane	0.409	0.419	-2.4	96	0.00
62 S	4-Bromofluorobenzene	0.484	0.494	-2.1	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.366	0.365	0.3	95	0.00
65 PM	Chlorobenzene	1.120	1.107	1.2	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.436	-3.3	99	0.00
67 C	Ethyl Benzene	2.003	2.023	-1.0#	95	0.00
68 T	m/p-Xylenes	0.778	0.785	-0.9	95	0.00
69 T	o-Xylene	0.756	0.767	-1.5	95	0.00
70 T	Styrene	1.257	1.291	-2.7	95	0.00
71 P	Bromoform	0.333	0.348	-4.5	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.962	3.825	3.5	95	0.00
74 T	N-amyl acetate	1.497	1.472	1.7	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.272	1.189	6.5	93	0.00
76 T	1,2,3-Trichloropropane	1.112	1.084	2.5	94	0.00
77 T	Bromobenzene	0.971	0.940	3.2	97	0.00
78 T	n-propylbenzene	4.565	4.414	3.3	94	0.00
79 T	2-Chlorotoluene	2.734	2.656	2.9	95	0.00
80 T	1,3,5-Trimethylbenzene	3.386	3.337	1.4	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.350	0.340	2.9	91	0.00
82 T	4-Chlorotoluene	3.199	3.135	2.0	95	0.00
83 T	tert-Butylbenzene	3.381	3.339	1.2	96	0.00
84 T	1,2,4-Trimethylbenzene	3.368	3.322	1.4	94	0.00
85 T	sec-Butylbenzene	4.068	4.029	1.0	94	0.00
86 T	p-Isopropyltoluene	3.452	3.447	0.1	93	0.00
87 T	1,3-Dichlorobenzene	1.832	1.794	2.1	96	0.00
88 T	1,4-Dichlorobenzene	1.901	1.809	4.8	95	0.00
89 T	n-Butylbenzene	3.003	2.963	1.3	92	0.00
90 T	Hexachloroethane	0.564	0.574	-1.8	96	0.00
91 T	1,2-Dichlorobenzene	1.807	1.768	2.2	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.268	3.9	91	0.00
93 T	1,2,4-Trichlorobenzene	1.178	1.103	6.4	94	0.00
94 T	Hexachlorobutadiene	0.492	0.454	7.7	93	0.00
95 T	Naphthalene	3.925	3.689	6.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103122\
Data File : VX032499.D
Acq On : 31 Oct 2022 19:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 01 05:17:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.197	1.097	8.4	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6