

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048417.D
 Acq On : 31 Oct 2025 09:46
 Operator : JC/MD
 Sample : VX1031MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1031MBL01

Quant Time: Oct 31 22:59:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	106852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	212289	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	217921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	109689	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	80634	52.472	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.940%
35) Dibromofluoromethane	5.367	113	61649	50.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.880%
50) Toluene-d8	8.629	98	260619	48.780	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.560%
62) 4-Bromofluorobenzene	11.061	95	112962	56.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048417.D
Acq On : 31 Oct 2025 09:46
Operator : JC/MD
Sample : VX1031MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1031MBL01

Quant Time: Oct 31 22:59:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

