

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048421.D
 Acq On : 31 Oct 2025 11:21
 Operator : JC/MD
 Sample : Q3480-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RT3449

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 11/03/2025
 Supervised By :Mahesh Dadoda 11/03/2025

Quant Time: Oct 31 23:02:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	108336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	214464	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	223837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	113437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	85153	54.654	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.300%
35) Dibromofluoromethane	5.373	113	63925	52.280	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.560%
50) Toluene-d8	8.635	98	270120	50.046	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.061	95	114466	56.692	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.380%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	5449	15.431	ug/l	96
20) Methylene Chloride	2.794	84	7484	5.498	ug/l	92
25) 2-Butanone	4.574	43	3657m	6.483	ug/l	
43) Isopropyl Acetate	6.330	43	9811	3.397	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048421.D
Acq On : 31 Oct 2025 11:21
Operator : JC/MD
Sample : Q3480-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RT3449

Quant Time: Oct 31 23:02:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 11/03/2025
Supervised By :Mahesh Dadoda 11/03/2025

