

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048423.D
 Acq On : 31 Oct 2025 12:02
 Operator : JC/MD
 Sample : Q3500-03 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 463

Quant Time: Oct 31 23:03:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	120387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	246735	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	100604	58.107	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.220%
35) Dibromofluoromethane	5.373	113	74641	53.060	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.120%
50) Toluene-d8	8.635	98	305059	49.127	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.061	95	134623	57.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.900%
Target Compounds						
16) Acetone	2.380	43	1771	4.513	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048423.D
Acq On : 31 Oct 2025 12:02
Operator : JC/MD
Sample : Q3500-03 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
463

Quant Time: Oct 31 23:03:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

