

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048426.D
 Acq On : 31 Oct 2025 13:04
 Operator : JC/MD
 Sample : Q3494-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Field

Quant Time: Oct 31 23:04:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	158807	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	321644	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	335954	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	168431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	124876	54.677	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.360%	
35) Dibromofluoromethane	5.379	113	93588	51.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.060%	
50) Toluene-d8	8.635	98	389545	48.122	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.240%	
62) 4-Bromofluorobenzene	11.061	95	172045	56.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.640%	
Target Compounds						
16) Acetone	2.380	43	2769	5.349	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048426.D
Acq On : 31 Oct 2025 13:04
Operator : JC/MD
Sample : Q3494-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Field

Quant Time: Oct 31 23:04:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

