

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048432.D
 Acq On : 31 Oct 2025 15:08
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 31 23:06:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	142473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	328670	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	340984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	174424	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	108913	53.155	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.300%
35) Dibromofluoromethane	5.373	113	80286	42.845	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.680%
50) Toluene-d8	8.635	98	404224	48.868	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.061	95	175220	56.627	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048432.D
Acq On : 31 Oct 2025 15:08
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 31 23:06:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

