

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
 Data File : VX048442.D
 Acq On : 31 Oct 2025 18:35
 Operator : JC/MD
 Sample : Q3511-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL-001

Quant Time: Oct 31 23:10:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 30 02:37:17 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	106505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	213921	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	262049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	138313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83058	54.226	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.460%	
35) Dibromofluoromethane	5.373	113	62773	51.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.940%	
50) Toluene-d8	8.634	98	304075	56.480	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 112.960%#	
62) 4-Bromofluorobenzene	11.061	95	136198	67.626	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 135.260%#	
Target Compounds						
16) Acetone	2.373	43	4196	12.087	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX103125\
Data File : VX048442.D
Acq On : 31 Oct 2025 18:35
Operator : JC/MD
Sample : Q3511-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL-001

Quant Time: Oct 31 23:10:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102925W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 30 02:37:17 2025
Response via : Initial Calibration

