

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005721.D
 Acq On : 01 Nov 2018 11:30
 Operator : JC/MD
 Sample : VX1101MBL01
 Misc : 5.0g/10mL/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1101MBL01

Quant Time: Nov 01 14:47:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	127552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	204526	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	173610	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	76958	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	93044	49.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.78%	
35) Dibromofluoromethane	5.50	113	72101	50.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.82%	
50) Toluene-d8	8.71	98	246846	49.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.40%	
62) 4-Bromofluorobenzene	11.14	95	79243	43.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005721.D
Acq On : 01 Nov 2018 11:30
Operator : JC/MD
Sample : VX1101MBL01
Misc : 5.0g/10mL/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1101MBL01

Quant Time: Nov 01 14:47:51 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

