

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005722.D
 Acq On : 01 Nov 2018 11:56
 Operator : JC/MD
 Sample : VX1101WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1101WBL01

Quant Time: Nov 01 14:52:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	150646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	233973	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	202304	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	91459	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	109306	49.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.26%	
35) Dibromofluoromethane	5.51	113	79118	48.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.68%	
50) Toluene-d8	8.72	98	288581	50.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.58%	
62) 4-Bromofluorobenzene	11.14	95	92767	44.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005722.D
Acq On : 01 Nov 2018 11:56
Operator : JC/MD
Sample : VX1101WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1101WBL01

Quant Time: Nov 01 14:52:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

