

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005724.D
 Acq On : 01 Nov 2018 12:57
 Operator : JC/MD
 Sample : J5199-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-01-103018

Quant Time: Nov 01 15:04:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	142322	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	244320	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	197670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	86459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	103578	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
35) Dibromofluoromethane	5.51	113	78268	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
50) Toluene-d8	8.71	98	276478	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
62) 4-Bromofluorobenzene	11.14	95	90226	41.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.40%	
Target Compounds						
18) Methyl Acetate	2.78	43	3126	1.047	ug/l	96
20) Methylene Chloride	2.85	84	2256	1.257	ug/l	94
43) Isopropyl Acetate	6.46	43	100333	20.816	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005724.D
Acq On : 01 Nov 2018 12:57
Operator : JC/MD
Sample : J5199-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-01-103018

Quant Time: Nov 01 15:04:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

