

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005729.D
 Acq On : 01 Nov 2018 15:11
 Operator : JC/MD
 Sample : J5769-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-10-31-18

Quant Time: Nov 01 16:22:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	118273	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	199880	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	163065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	74051	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	89841	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
35) Dibromofluoromethane	5.50	113	64667	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
50) Toluene-d8	8.71	98	229957	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
62) 4-Bromofluorobenzene	11.14	95	75181	41.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.94%	
Target Compounds						
20) Methylene Chloride	2.86	84	1603	1.075	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005729.D
Acq On : 01 Nov 2018 15:11
Operator : JC/MD
Sample : J5769-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-10-31-18

Quant Time: Nov 01 16:22:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

