

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005730.D
 Acq On : 01 Nov 2018 15:37
 Operator : JC/MD
 Sample : J5769-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-10-31-18

Quant Time: Nov 01 16:24:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	128266	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	201813	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	180393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	79005	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	94596	50.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.88%	
35) Dibromofluoromethane	5.50	113	72161	51.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.28%	
50) Toluene-d8	8.71	98	250815	51.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.14	95	80151	44.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005730.D
Acq On : 01 Nov 2018 15:37
Operator : JC/MD
Sample : J5769-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-10-31-18

Quant Time: Nov 01 16:24:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

