

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005732.D
 Acq On : 01 Nov 2018 16:31
 Operator : JC/MD
 Sample : J5729-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WS1

Quant Time: Nov 02 12:35:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	115468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	194727	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	176369	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	80406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	89547	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
35) Dibromofluoromethane	5.50	113	67609	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
50) Toluene-d8	8.72	98	246329	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	11.14	95	82164	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	20834	45.444	ug/l	98
16) Acetone	2.45	43	34007	31.014	ug/l	96
52) Toluene	8.79	92	6783	2.032	ug/l	97
67) Ethyl Benzene	10.26	91	2585	0.404	ug/l	94
68) m/p-Xylenes	10.36	106	3267	1.375	ug/l	94
69) o-Xylene	10.70	106	1557	0.685	ug/l	98
84) 1,2,4-Trimethylbenzene	11.81	105	2178	0.479	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005732.D
Acq On : 01 Nov 2018 16:31
Operator : JC/MD
Sample : J5729-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WS1

Quant Time: Nov 02 12:35:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

