

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
 Data File : VX005737.D
 Acq On : 01 Nov 2018 18:45
 Operator : JC/MD
 Sample : PB114370ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114370ZHE#03

Quant Time: Nov 02 04:25:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 31 18:23:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	136616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	205215	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	181871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	78159	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	98573	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	5.51	113	69165	48.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.36%	
50) Toluene-d8	8.71	98	253981	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
62) 4-Bromofluorobenzene	11.14	95	81619	44.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.74%	
Target Compounds						
16) Acetone	2.45	43	11145	1.592	ug/l	91
18) Methyl Acetate	2.78	43	7210	2.515	ug/l	97
20) Methylene Chloride	2.85	84	4952	2.875	ug/l	93
25) 2-Butanone	4.71	43	2059	1.222	ug/l	99
43) Isopropyl Acetate	6.46	43	104296	25.761	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX110118\
Data File : VX005737.D
Acq On : 01 Nov 2018 18:45
Operator : JC/MD
Sample : PB114370ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114370ZHE#03

Quant Time: Nov 02 04:25:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103118W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 31 18:23:06 2018
Response via : Initial Calibration

